

SOL # 241
ARWF-0394

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01582	Calib. Date:	04/18/2017
Certification File No.:	01583	Cert. No.:	00037
Linearity File No.:	01584	Lin. Date:	04/18/2017
Solution File No.:	01601	Cert. No.:	00027
Sequential File No.:	01601	Lin. No.:	00027
		Soln. Date:	05/12/2017
		Soln. No.:	00241
		File Date:	05/12/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1283


Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:22D		
Control 1 EC	0.100%	15:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	15:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:24D		
Control 2 EC	0.099%	15:24D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	15:24D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:25D		
Control 3 EC	0.099%	15:26D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:26D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN	First Name: ADAM	MI: M
--------------------	------------------	-------

Signature:  # 350	Badge No.: 350
	Date: 05/12/2017