

SOL # 242

6-3-2017


Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02054	Calib. Date:	05/12/2017
Certification File No.:	02055	Cert. Date:	05/12/2017
Linearity File No.:	02056	Lin. Date:	05/12/2017
Solution File No.:	02078	Soln. Date:	06/03/2017
Sequential File No.:	02078	File Date:	06/03/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	16390	Expires:	12/02/2018
		Bottle No.:	1343

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	00:48D		
Control 1 EC	0.099%	00:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	00:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:49D		
Control 2 EC	0.099%	00:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	00:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:50D		
Control 3 EC	0.099%	00:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:51D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:
 Last Name: GAMBARDELLA First Name: PASCAL MI: M
 Signature:  #365 Badge No.: 365
 Date: 06/03/2017