

Calibrating Unit New Standard Solution Report



ORIGINAL

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02054
Certification File No.: 02055
Linearity File No.: 02056
Solution File No.: 02097
Sequential File No.: 02097

Serial No.: ARAJ-0075

Calib. Date: 05/12/2017
Cert. Date: 05/12/2017
Lin. Date: 05/12/2017
Soln. Date: 06/25/2017
File Date: 06/25/2017

Calib. No.: 00028
Cert. No.: 00019
Lin. No.: 00019
Soln. No.: 00244

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 03/20/2019
Bottle No.: 0226

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:46D		
Control 1 EC	0.099%	06:47D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:47D		
Control 2 EC	0.099%	06:48D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:49D		
Control 3 EC	0.100%	06:49D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:49D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:50D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 06/25/2017