

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01582 Calib. Date: 04/18/2017 Calib. No.: 00037
Certification File No.: 01583 Cert. Date: 04/18/2017 Cert. No.: 00027
Linearity File No.: 01584 Lin. Date: 04/18/2017 Lin. No.: 00027
Solution File No.: 01610 Soln. Date: 06/30/2017 Soln. No.: 00243
Sequential File No.: 01610 File Date: 06/30/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387
Control Solution %: 0.100% Expires: 03/20/2019
Solution Control Lot: 17110 Bottle No.: 0601

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	15:42D		
Control 1 EC	0.099%	15:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:42D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:43D		
Control 2 EC	0.098%	15:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44D		
Control 3 EC	0.099%	15:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:46D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 06/30/2017