

SOL # 245  
7/17/17

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02054  
Certification File No.: 02055  
Linearity File No.: 02056  
Solution File No.: 02113  
Sequential File No.: 02113

Serial No.: ARAJ-0075  
Calib. Date: 05/12/2017  
Cert. Date: 05/12/2017  
Lin. Date: 05/12/2017  
Soln. Date: 07/17/2017  
File Date: 07/17/2017

Calib. No.: 00028  
Cert. No.: 00019  
Lin. No.: 00019  
Soln. No.: 00245

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 16390

Model No.: CU-34

Serial No.: DDWK S3-0386  
Expires: 12/02/2018  
Bottle No.: 1271

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:00D		
Control 1 EC	0.099%	15:01D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:01D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:02D		
Control 2 EC	0.099%	15:02D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:02D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:03D		
Control 3 EC	0.100%	15:03D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:03D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:04D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: NATER

First Name: MARK

MI: S

Signature: PH. J. Nater #336

Badge No.: 336

Date: 07/17/2017

 ORIGINAL