

SOLN # 246

7/30/17

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02054
Certification File No.: 02055
Linearity File No.: 02056
Solution File No.: 02120
Sequential File No.: 02120

Serial No.: ARAJ-0075
Calib. No.: 00028
Cert. No.: 00019
Lin. No.: 00019
Soln. No.: 00246
File Date: 07/30/2017

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 16390

Model No.: CU-34
Serial No.: DDWK S3-0386
Expires: 12/02/2018
Bottle No.: 1214

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	06:13D		
Control 1 EC	0.099%	06:14D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	06:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:15D		
Control 2 EC	0.098%	06:15D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	06:15D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:16D		
Control 3 EC	0.099%	06:16D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	06:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:17D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WORTH

First Name: ADAM

MI: C

Signature: 

Badge No.: 378

Date: 07/30/2017