

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS RIVER P.D.	
Calibration File No.:	01582	Calib. Date: 04/18/2017
Certification File No.:	01583	Calib. No.: 00037
Linearity File No.:	01584	Cert. Date: 04/18/2017
Solution File No.:	01619	Cert. No.: 00027
Sequential File No.:	01619	Lin. Date: 04/18/2017
		Lin. No.: 00027
		Soln. Date: 08/19/2017
		Soln. No.: 00245
		File Date: 08/19/2017
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	16390	Expires: 12/02/2018
		Bottle No.: 1212

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:53D		
Control 1 EC	0.098%	04:53D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:53D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:54D		
Control 2 EC	0.097%	04:55D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	04:55D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:55D		
Control 3 EC	0.097%	04:56D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:56D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:57D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DEROSA

First Name: MICHAEL

MI: A

Signature: _____

Badge No.: 386

Date: 08/19/2017