

8/28/17  
# 248



# ORIGINAL **Calibrating Unit** New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
 Location: TOMS RIVER P.D.  
 Calibration File No.: 02054 Calib. Date: 05/12/2017 Calib. No.: 00028  
 Certification File No.: 02055 Cert. Date: 05/12/2017 Cert. No.: 00019  
 Linearity File No.: 02056 Lin. Date: 05/12/2017 Lin. No.: 00019  
 Solution File No.: 02134 Soln. Date: 08/28/2017 Soln. No.: 00248  
 Sequential File No.: 02134 File Date: 08/28/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386  
 Control Solution %: 0.100% Expires: 12/02/2018  
 Solution Control Lot: 16390 Bottle No.: 1341

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	21:38D		
Control 1 EC	0.098%	21:39D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	21:39D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:39D		
Control 2 EC	0.098%	21:40D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:40D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40D		
Control 3 EC	0.099%	21:41D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:41D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:42D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: \_\_\_\_\_

Badge No.: 306

Date: 08/28/2017