

SOL # 249

9/1/2017

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-----------------------|--------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARAJ-0075 |
| Location: | TOMS RIVER P.D. | | |
| Calibration File No.: | 02054 | Calib. Date: | 05/12/2017 |
| Certification File No.: | 02055 | Cert. No.: | 00028 |
| Linearity File No.: | 02056 | Cert. Date: | 05/12/2017 |
| Solution File No.: | 02137 | Lin. No.: | 00019 |
| Sequential File No.: | 02137 | Lin. Date: | 05/12/2017 |
| | | Soln. Date: | 09/01/2017 |
| | | Soln. No.: | 00249 |
| | | File Date: | 09/01/2017 |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWK S3-0386 |
| Solution Control Lot: | 16390 | Expires: | 12/02/2018 |
| | | Bottle No.: | 1336 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 08:30D | | |
| Control 1 EC | 0.099% | 08:31D | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.100% | 08:31D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:31D | | |
| Control 2 EC | 0.099% | 08:32D | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 08:32D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:33D | | |
| Control 3 EC | 0.099% | 08:33D | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.100% | 08:33D | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 08:34D | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WORTH

First Name: ADAM

MI: C

Signature: _____



Badge No.: 378

Date: 09/01/2017