

SOL # 250  
9/8/2017

# Calibrating Unit New Standard Solution Report


**Equipment** Alcotest 7110 MKIII-C Serial No.: ARAJ-0075  
Location: TOMS RIVER P.D.  
Calibration File No.: 02054 Calib. Date: 05/12/2017 Calib. No.: 00028  
Certification File No.: 02055 Cert. Date: 05/12/2017 Cert. No.: 00019  
Linearity File No.: 02056 Lin. Date: 05/12/2017 Lin. No.: 00019  
Solution File No.: 02144 Soln. Date: 09/08/2017 Soln. No.: 00250  
Sequential File No.: 02144 File Date: 09/08/2017

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0386  
Control Solution %: 0.100% Expires: 03/20/2019  
Solution Control Lot: 17110 Bottle No.: 0291

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 05:59D        |                               |                           |
| Control 1 EC      | -.---%         | 05:59D        | 34.0°C                        | CTRL GAS SUPPLY           |
| Control 1 IR      | -.---%         | 05:59D        | 34.0°C                        | CTRL GAS SUPPLY           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |
| Control 2 EC      | -.---%         | --:--         | N/A                           |                           |
| Control 2 IR      | -.---%         | --:--         | N/A                           |                           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |
| Control 3 EC      | -.---%         | --:--         | N/A                           |                           |
| Control 3 IR      | -.---%         | --:--         | N/A                           |                           |
| Ambient Air Blank | -.---%         | --:--         |                               |                           |

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**  
Last Name: DEROSA First Name: MICHAEL MI: A  
Signature:  Badge No.: 386  
Date: 09/08/2017