

Calibrating Unit

New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D.
 Calibration File No.: 01631
 Certification File No.: 01632
 Linearity File No.: 01633
 Solution File No.: 01634
 Sequential File No.: 01634

Serial No.: ARWF-0394

Calib. Date: 09/15/2017 Calib. No.: 00038
 Cert. Date: 09/15/2017 Cert. No.: 00028
 Lin. Date: 09/15/2017 Lin. No.: 00028
 Soln. Date: 09/15/2017 Soln. No.: 00247
 File Date: 09/15/2017

Calibrating Unit: WET
 Control Solution %: 0.100%
 Solution Control Lot: 17110

Model No.: CU-34 Serial No.: DDWK S3-0387
 Expires: 03/20/2019
 Bottle No.: 0618

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	13:11D		
Control 1 EC	0.101%	13:12D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	13:12D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:13D		
Control 2 EC	0.100%	13:13D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	13:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:14D		
Control 3 EC	0.100%	13:14D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	13:14D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	13:15D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJPZ-026 (KA)

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpt. I K W. Q #6704

Badge No.: 6704

Date: 09/15/2017