

ITEM 14

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02147
Certification File No.: 02148
Linearity File No.: 02149
Solution File No.: 02150
Sequential File No.: 02150

Serial No.: ARAJ-0075

Calib. Date: 09/15/2017 Calib. No.: 00029
Cert. Date: 09/15/2017 Cert. No.: 00020
Lin. Date: 09/15/2017 Lin. No.: 00020
Soln. Date: 09/15/2017 Soln. No.: 00252
File Date: 09/15/2017

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110

Model No.: CU-34 Serial No.: DDWK S3-0386
Expires: 03/20/2019
Bottle No.: 0495

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:14D		
Control 1 EC	0.102%	11:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	11:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:15D		
Control 2 EC	0.101%	11:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:17D		
Control 3 EC	0.101%	11:17D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	11:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ PZ - 025 (KA)

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: TP. I MK W. Q #6704

Badge No.: 6704

Date: 09/15/2017