

SOL # 248

10/14/17

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01631
Certification File No.: 01632
Linearity File No.: 01633
Solution File No.: 01636
Sequential File No.: 01636

Serial No.: ARWF-0394
Calib. No.: 00038
Cert. No.: 00028
Lin. No.: 00028
Soln. No.: 00248
File Date: 10/14/2017

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110

Model No.: CU-34
Serial No.: DDWK S3-0387
Expires: 03/20/2019
Bottle No.: 0610

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	00:26D		
Control 1 EC	0.100%	00:27D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	00:27D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:27D		
Control 2 EC	0.099%	00:28D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	00:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:29D		
Control 3 EC	0.100%	00:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	00:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:30D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 10/14/2017