

SOLN # 255
11/4/17

Calibrating Unit New Standard Solution Report

| | | | |
|-------------------------|-----------------------|-------------|--------------|
| Equipment | Alcotest 7110 MKIII-C | Serial No.: | ARAJ-0075 |
| Location: | TOMS RIVER P.D. | Calib. No.: | 00029 |
| Calibration File No.: | 02147 | Cert. Date: | 09/15/2017 |
| Certification File No.: | 02148 | Lin. Date: | 09/15/2017 |
| Linearity File No.: | 02149 | Soln. Date: | 11/04/2017 |
| Solution File No.: | 02177 | File Date: | 11/04/2017 |
| Sequential File No.: | 02177 | | |
| | | | |
| Calibrating Unit: | WET | Model No.: | CU-34 |
| Control Solution %: | 0.100% | Serial No.: | DDWK S3-0386 |
| Solution Control Lot: | 17110 | Expires: | 03/20/2019 |
| | | Bottle No.: | 0216 |

| Function | Result %BAC | Time HH:MM | Temperature Simulator (°C) | Comment(s) or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000% | 05:14S | | |
| Control 1 EC | 0.099% | 05:14S | 34.0°C | *** TEST PASSED *** |
| Control 1 IR | 0.099% | 05:14S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 05:15S | | |
| Control 2 EC | 0.099% | 05:16S | 34.0°C | *** TEST PASSED *** |
| Control 2 IR | 0.100% | 05:16S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 05:16S | | |
| Control 3 EC | 0.099% | 05:17S | 34.0°C | *** TEST PASSED *** |
| Control 3 IR | 0.099% | 05:17S | 34.0°C | *** TEST PASSED *** |
| Ambient Air Blank | 0.000% | 05:17S | | |

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

| | | |
|--------------------|---------------------|------------------|
| Last Name: DE ROSA | First Name: MICHAEL | MI: A |
| Signature: | | Badge No.: 386 |
| | | Date: 11/04/2017 |