

SOL# 256

Calibrating Unit New Standard Solution Report

ORIGINAL 11/6/17

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02147
Certification File No.: 02148
Linearity File No.: 02149
Solution File No.: 02180
Sequential File No.: 02180

Calib. Date: 09/15/2017
Cert. Date: 09/15/2017
Lin. Date: 09/15/2017
Soln. Date: 11/06/2017
File Date: 11/06/2017

Serial No.: ARAJ-0075
Calib. No.: 00029
Cert. No.: 00020
Lin. No.: 00020
Soln. No.: 00256

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 03/20/2019
Bottle No.: 0603

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:18S		
Control 1 EC	0.097%	16:19S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	16:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:19S		
Control 2 EC	0.097%	16:20S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	16:20S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:20S		
Control 3 EC	0.098%	16:21S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:21S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:22S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature: _____

[Handwritten Signature] # 306

Badge No.: 306

Date: 11/06/2017