

SOL # 250

11-22-17

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01631	Calib. Date:	09/15/2017
Certification File No.:	01632	Cert. No.:	00038
Linearity File No.:	01633	Cert. Date:	09/15/2017
Solution File No.:	01645	Lin. No.:	00028
Sequential File No.:	01645	Lin. Date:	09/15/2017
		Soln. Date:	11/22/2017
		Soln. No.:	00250
		File Date:	11/22/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	17110	Expires:	03/20/2019
		Bottle No.:	0089

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	09:03S		
Control 1 EC	0.100%	09:04S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	09:04S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:04S		
Control 2 EC	0.099%	09:05S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	09:05S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:06S		
Control 3 EC	0.099%	09:07S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	09:07S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	09:07S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKPZ-444 (KA)

**Changed By:**

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpr. I K W. #6704

Badge No.: 6704

Date: 11/22/2017