

SOL # 251

12/14/17

# Calibrating Unit New Standard Solution Report

|                         |                       |             |            |
|-------------------------|-----------------------|-------------|------------|
| <b>Equipment</b>        | Alcotest 7110 MKIII-C | Serial No.: | ARWF-0394  |
| Location:               | TOMS RIVER P.D.       | Calib. No.: | 00038      |
| Calibration File No.:   | 01631                 | Cert. Date: | 09/15/2017 |
| Certification File No.: | 01632                 | Lin. Date:  | 09/15/2017 |
| Linearity File No.:     | 01633                 | Soln. Date: | 12/14/2017 |
| Solution File No.:      | 01649                 | File Date:  | 12/14/2017 |
| Sequential File No.:    | 01649                 |             |            |
| <br>                    |                       |             |            |
| Calibrating Unit:       | WET                   | Model No.:  | CU-34      |
| Control Solution %:     | 0.100%                | Serial No.: | DDAN-0040  |
| Solution Control Lot:   | 17290                 | Expires:    | 09/14/2019 |
|                         |                       | Bottle No.: | 1479       |

| Function          | Result<br>%BAC | Time<br>HH:MM | Temperature<br>Simulator (°C) | Comment(s)<br>or Error(s) |
|-------------------|----------------|---------------|-------------------------------|---------------------------|
| Ambient Air Blank | 0.000%         | 11:31S        |                               |                           |
| Control 1 EC      | 0.100%         | 11:32S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 1 IR      | 0.100%         | 11:32S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:33S        |                               |                           |
| Control 2 EC      | 0.098%         | 11:33S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 2 IR      | 0.097%         | 11:33S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:34S        |                               |                           |
| Control 3 EC      | 0.098%         | 11:35S        | 33.9°C                        | *** TEST PASSED ***       |
| Control 3 IR      | 0.098%         | 11:35S        | 33.9°C                        | *** TEST PASSED ***       |
| Ambient Air Blank | 0.000%         | 11:36S        |                               |                           |

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 12/14/2017