

12/29/2017

SOL # 259

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.	Calib. Date:	09/15/2017
Calibration File No.:	02147	Calib. No.:	00029
Certification File No.:	02148	Cert. Date:	09/15/2017
Linearity File No.:	02149	Lin. Date:	09/15/2017
Solution File No.:	02230	Soln. Date:	12/29/2017
Sequential File No.:	02230	File Date:	12/29/2017
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	1472

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:49S		
Control 1 EC	0.098%	14:49S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:49S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:50S		
Control 2 EC	0.098%	14:50S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:50S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51S		
Control 3 EC	0.098%	14:52S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:53S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA First Name: PASCAL MI: M

Signature:  #365 Badge No.: 365 Date: 12/29/2017