

1/2/2018

# Calibrating Unit New Standard Solution Report

SOL # 260

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 02147  
Certification File No.: 02148  
Linearity File No.: 02149  
Solution File No.: 02234  
Sequential File No.: 02234

Calib. Date: 09/15/2017  
Cert. Date: 09/15/2017  
Lin. Date: 09/15/2017  
Soln. Date: 01/02/2018  
File Date: 01/02/2018

Serial No.: ARAJ-0075  
Calib. No.: 00029  
Cert. No.: 00020  
Lin. No.: 00020  
Soln. No.: 00260

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 17110

Model No.: CU-34

Serial No.: DDAN-0041  
Expires: 03/20/2019  
Bottle No.: 0211

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:29S		
Control 1 EC	0.097%	12:30S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	12:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:31S		
Control 2 EC	0.096%	12:31S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	12:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:32S		
Control 3 EC	0.096%	12:33S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	12:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:33S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: SCHWARTZ

First Name: STEVE

MI: C

Signature: \_\_\_\_\_



28

Badge No.: 1293

Date: 01/02/2018