

SOL # 261

1/7/2018

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 02147
Certification File No.: 02148
Linearity File No.: 02149
Solution File No.: 02238
Sequential File No.: 02238

Serial No.: ARAJ-0075

Calib. Date: 09/15/2017
Cert. Date: 09/15/2017
Lin. Date: 09/15/2017
Soln. Date: 01/07/2018
File Date: 01/07/2018

Calib. No.: 00029
Cert. No.: 00020
Lin. No.: 00020
Soln. No.: 00261

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17290

Model No.: CU-34

Serial No.: DDAN-0041
Expires: 09/14/2019
Bottle No.: 1240

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:31S		
Control 1 EC	0.097%	08:31S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	08:31S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:32S		
Control 2 EC	0.095%	08:33S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.097%	08:33S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:34S		
Control 3 EC	0.096%	08:34S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	08:34S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:35S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTFALL

First Name: ROBERT

MI: S

Signature:  FJ 306

Badge No.: 306

Date: 01/07/2018