

# Calibrating Unit New Standard Solution Report

SOLN 253  
1/26/18

**Equipment** Alcotest 7110 MKIII-C  
Location: TOMS RIVER P.D.  
Calibration File No.: 01631  
Certification File No.: 01632  
Linearity File No.: 01633  
Solution File No.: 01658  
Sequential File No.: 01658

Serial No.: ARWF-0394

Calib. Date: 09/15/2017 Calib. No.: 00038  
Cert. Date: 09/15/2017 Cert. No.: 00028  
Lin. Date: 09/15/2017 Lin. No.: 00028  
Soln. Date: 01/26/2018 Soln. No.: 00253  
File Date: 01/26/2018

Calibrating Unit: WET  
Control Solution %: 0.100%  
Solution Control Lot: 17110

Model No.: CU-34

Serial No.: DDAN-0040  
Expires: 03/20/2019  
Bottle No.: 0090

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:35S		
Control 1 EC	0.101%	15:36S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	15:36S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:37S		
Control 2 EC	0.100%	15:37S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:37S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:38S		
Control 3 EC	0.100%	15:39S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	15:39S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:40S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

## Changed By:

Last Name: GAMBARADELLA

First Name: PASCAL

MI: M

Signature: 

Badge No.: 365

Date: 01/26/2018