

SOLN 263
1/26/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C Serial No.: ARAJ-0075
Location: TOMS RIVER P.D.
Calibration File No.: 02147 Calib. Date: 09/15/2017 Calib. No.: 00029
Certification File No.: 02148 Cert. Date: 09/15/2017 Cert. No.: 00020
Linearity File No.: 02149 Lin. Date: 09/15/2017 Lin. No.: 00020
Solution File No.: 02250 Soln. Date: 01/26/2018 Soln. No.: 00263
Sequential File No.: 02250 File Date: 01/26/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0041
Control Solution %: 0.100% Expires: 09/14/2019
Solution Control Lot: 17290 Bottle No.: 1301

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:40S		
Control 1 EC	0.100%	15:41S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	15:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:41S		
Control 2 EC	0.099%	15:42S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	15:42S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:43S		
Control 3 EC	0.099%	15:43S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	15:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:44S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 01/26/2018