

Item 14

Calibrating Unit New Standard Solution Report

SOLN 254
1-30-18

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01659
Certification File No.: 01660
Linearity File No.: 01661
Solution File No.: 01662
Sequential File No.: 01662

Serial No.: ARWF-0394

Calib. Date: 01/30/2018 Calib. No.: 00039
Cert. Date: 01/30/2018 Cert. No.: 00029
Lin. Date: 01/30/2018 Lin. No.: 00029
Soln. Date: 01/30/2018 Soln. No.: 00254
File Date: 01/30/2018

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17110

Model No.: CU-34 Serial No.: DDAN-0040
Expires: 03/20/2019
Bottle No.: 0665

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:41S		
Control 1 EC	0.101%	14:41S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	14:41S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:42S		
Control 2 EC	0.101%	14:43S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	14:43S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:44S		
Control 3 EC	0.101%	14:45S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	14:45S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DXKP2-444 (KA)

Changed By:

Last Name: ALCOTT

First Name: KEVIN

MI: W.

Signature: Tpt. I K W. Q #6704

Badge No.: 6704

Date: 01/30/2018