

SOLN 255
2/16/18

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01659	Calib. Date:	01/30/2018
Certification File No.:	01660	Cert. No.:	00039
Linearity File No.:	01661	Lin. Date:	01/30/2018
Solution File No.:	01664	Lin. No.:	00029
Sequential File No.:	01664	Soln. Date:	02/16/2018
		Soln. No.:	00255
		File Date:	02/16/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	1485

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:26S		
Control 1 EC	0.099%	12:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27S		
Control 2 EC	0.099%	12:28S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:28S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29S		
Control 3 EC	0.099%	12:30S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:30S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA First Name: PASCAL MI: M

Signature: *Pascal Gambardella #365*

Badge No.: 365
Date: 02/16/2018