

SOLN # 256

Calibrating Unit New Standard Solution Report

3/8/18

Equipment Alcotest 7110 MKIII-C
 Location: TOMS RIVER P.D. Serial No.: ARWF-0394
 Calibration File No.: 01659 Calib. Date: 01/30/2018 Calib. No.: 00039
 Certification File No.: 01660 Cert. Date: 01/30/2018 Cert. No.: 00029
 Linearity File No.: 01661 Lin. Date: 01/30/2018 Lin. No.: 00029
 Solution File No.: 01667 Soln. Date: 03/08/2018 Soln. No.: 00256
 Sequential File No.: 01667 File Date: 03/08/2018

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
 Control Solution %: 0.100% Expires: 09/14/2019
 Solution Control Lot: 17290 Bottle No.: 1371

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	16:53S		
Control 1 EC	0.101%	16:54S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	16:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:55S		
Control 2 EC	0.100%	16:55S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:55S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:56S		
Control 3 EC	0.100%	16:57S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	16:57S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:58S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 03/08/2018