

SOLN # 257

Calibrating Unit New Standard Solution Report

3/29/18

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01659	Calib. Date:	01/30/2018
Certification File No.:	01660	Cert. Date:	01/30/2018
Linearity File No.:	01661	Lin. Date:	01/30/2018
Solution File No.:	01669	Soln. Date:	03/29/2018
Sequential File No.:	01669	File Date:	03/29/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0040
Solution Control Lot:	17290	Expires:	09/14/2019
		Bottle No.:	1287

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	11:15S		
Control 1 EC	0.100%	11:16S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	11:16S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:16S		
Control 2 EC	0.099%	11:17S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	11:17S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:18S		
Control 3 EC	0.100%	11:19S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	11:19S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:19S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 03/29/2018