

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02255	Calib. Date: 01/30/2018
Certification File No.:	02256	Calib. No.: 00030
Linearity File No.:	02257	Cert. Date: 01/30/2018
Solution File No.:	02363	Cert. No.: 00021
Sequential File No.:	02363	Lin. Date: 01/30/2018
		Lin. No.: 00021
		Soln. Date: 05/31/2018
		Soln. No.: 00277
		File Date: 05/31/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	17290	Expires: 09/14/2019
		Bottle No.: 1488

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:06D		
Control 1 EC	0.100%	12:06D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:06D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:07D		
Control 2 EC	0.099%	12:07D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:07D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:08D		
Control 3 EC	0.099%	12:09D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	12:09D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:09D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_

Badge No.: 365

Date: 05/31/2018