

#4 10/14/16
ARAJ-0075

STATE OF NEW JERSEY
DEPARTMENT OF LAW AND PUBLIC SAFETY
DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Toms River Police Department 255 Oak Ave Toms River, NJ 08753	2. Contact: Sgt. Robert Rankin	4. Date: 8/24/2016
	3. Phone Number: 732-349-0150	5. County: Ocean

6. Alcotest Instrument Serial Number: ARAJ-0075	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A
---	---	---

9. Reason for Service:
Complaint received of the instrument experiencing multiple control test failures after fuel cell was recently replaced.

10. Comments:
Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed control test failures due to low electrochemical sensor (EC) results. The instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.

Last Known Sequential File #: 1931

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Kevin W. Alcott #6704
Name & Badge Number (Print)

Tpr. I K W. A #6704 8/24/16
Signature Date

10/14/16

ARAJ-0075

STATE OF NEW JERSEY
 DEPARTMENT OF LAW AND PUBLIC SAFETY
 DIVISION OF STATE POLICE
BREATH TESTING INSTRUMENTATION SERVICE REPORT

1. Department: Toms River Police Department 255 Oak Ave Toms River, NJ 08753	2. Contact: Sgt. Robert Rankin	4. Date: 10/14/2016
	3. Phone Number: 732-349-0150	5. County: Ocean

6. Alcotest Instrument Serial Number: ARAJ-0075	7. Simulator Component Serial Number: N/A	8. Temperature Probe Component Serial Number: N/A
---	---	---

9. Reason for Service:
The alcotest instrument was returned from outside evaluation and placed back in service.

10. Comments:
See Draeger Return and Repair form.

Last Known Sequential File #: N/A

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Kevin W. Alcott #6704	<i>Tpr. I K W. A #6704</i>	10/14/16
Name & Badge Number (Print)	Signature	Date

10/14/16



RETURN AND REPAIR FORM

ARAJ-0075

Customer Information

B:

S:

Company Name: Toms River P.D., Toms River, NJ

Date Received: 08/31/2016

Date given to service: 08/31/2016

Carrier: FedEx UPS USPS

Shipping Method: GRD 3DAY 2DAY
 NDA-PRI NDA-STD

Product: 6510 6810
 7110 8610
 7510 DT5000
 7410 Upper-half 9510
 7410 Whole

Serial Number: AR AJ-0075

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description: A B Plus Demo Screener Trade In

Accessories

110V A/C Adapter Regulator Printer Ribbon Printer Paper
 Mouthpieces 9510 Stylus 9510 Top Cover Carrying Case
 Dry Gas Other (please specify) _____

Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W

Repair Notes: Recalibrated unit, performed and passed all QC.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 09/30/2016

10/14/16

ARAJ-0075

Packing Slip

Customer no Order No Order date
150059761 10854913 10/03/2016

Packing slip no. Ship date
81032362 10/03/2016

Please reference on inquiries

Ship to
150059763
TOWNSHIP OF TOMS RIVER POLICE DEPAR
ATTN: SGT. ROBERT RANKIN
OCEAN COUNTY
255 OAK AVENUE
TOMS RIVER, NJ 08753-3383
USA

Payer

Your Purchase Order Number

SGT. ROBERT RANKIN
Any questions? Please contact:
ROSEMARIE HUGHES

Customer
TOWNSHIP OF TOMS RIVER
OCEAN COUNTY
255 OAK AVE
TOMS RIVER, NJ 08753-0876
USA

Your sales person

CHRISTOPHER FAIRCHILD
Phone: 866-385-5900
Fax: 972-929-1260

Delivery terms
FA
FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description
		Ship via: Fed Ex Saver
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.
		7110:ARAJ-0075 REC'D 8/30/2016
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 EA

Draeger Safety Diagnostics Inc.
Accounting Address:
101 Technology Drive
Pittsburgh, PA 15275
Tel: (412) 788-5537
Fax: (412) 788-5598

Remit Wire Transfers To:
Citizens Bank
Acct. # 6209426615
Acct. Name: Draeger Safety Diagnostics Inc.
Transit Number: 036076150
Federal ID Number: 84-1600159

REMIT TO:
Draeger Safety Diagnostics Inc.
P.O. Box 536410
Pittsburgh, PA 15253-5906