

# 5 4/18/17  
 ARAJ-0075

STATE OF NEW JERSEY  
 DEPARTMENT OF LAW AND PUBLIC SAFETY  
 DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Police Department 255 Oak Ave Toms River, NJ 08753</b>		2. Contact: <b>Sgt. Robert Rankin</b>	4. Date: <b>3/3/2017</b>
		3. Phone Number: <b>732-349-0150</b>	5. County: <b>Ocean</b>

6. Alcotest Instrument Serial Number: <b>ARAJ-0075</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>
---	---	---

9. Reason for Service:  
**Complaint received of the instrument experiencing multiple control test failures.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed control test failures due to low electrochemical sensor (EC) results and low infrared sensor (IR) results. The instrument was placed out of service and returned to Draeger Safety Diagnostics for evaluation.**

Last Known Sequential File #: 2039

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Kevin W. Alcott #6704  
 Name & Badge Number (Print)

Tpr. I R. W. Q #6704 3/3/17  
 Signature Date

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Police Department</b> <b>255 Oak Ave</b> <b>Toms River, NJ 08753</b>	2. Contact: <b>Sgt. Robert Rankin</b> 3. Phone Number: <b>732-349-0150</b>	4. Date: <b>4/18/2017</b> 5. County: <b>Ocean</b>
--	---	--

6. Alcotest Instrument Serial Number: <b>ARAJ-0075</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>
---	---	---


9. Reason for Service:  
**The alcotest instrument was returned from outside evaluation and placed back in service.**

10. Comments:  
**See Draeger Return and Repair form.**

Last Known Sequential File #: N/A

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Kevin W. Alcott #6704 <small>Name &amp; Badge Number (Print)</small>	 <small>Signature</small>	4/18/17 <small>Date</small>
--	--	--------------------------------

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Toms River P.D., Toms River, NJ

Date Received: 03/22/2017

Date given to service: 03/22/2017

Carrier:  FedEx  UPS  USPS

Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR AJ-0075

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	
4414166	Labor	.5	
6808455	Fuel Cell	1	

Repair Notes: Replaced defective fuel cell.

Fuel Cell = ARJH-1417

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 03/31/17

# Packing Slip

Customer no    Order No    Order date  
 150059761    10860913    04/01/2017

Packing slip no.    Ship date  
 81038784    04/03/2017

Please reference on inquiries

Ship to  
 150059763  
 TOWNSHIP OF TOMS RIVER POLICE DEPAR  
 ATTN: SGT. ROBERT RANKIN  
 OCEAN COUNTY  
 255 OAK AVENUE  
 TOMS RIVER, NJ    08753-3383  
 USA

Payer

Your Purchase Order Number  
 04/01/2017  
 SGT. ROBERT RANKIN  
 Any questions? Please contact:  
 ROSEMARIE HUGHES

Customer  
 TOWNSHIP OF TOMS RIVER  
 OCEAN COUNTY  
 255 OAK AVE  
 TOMS RIVER, NJ    08753-0876  
 USA

Your sales person  
 CHRISTOPHER FAIRCHILD  
 Phone: 866-385-5900  
 Fax: 972-929-1260

Delivery terms  
 FA  
 FREIGHT ALLOWED

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110:ARAJ-0075 REC'D 3/21/17			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 1 / 0 EA			
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 1 / 0 EA			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 3135 Quarry Road  
 Telford, PA 18969 USA  
 Toll-free: +1-800-437-2437  
 Fax: +1-215-721-5410

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 536410  
 Pittsburgh, PA 15253-5906

# Packing Slip

Customer no    Order No    Order date  
150059761    10860913    04/01/2017

Packing slip no.    Ship date  
81038784    04/03/2017  
Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	0.5 EA	4414166 HOURLY LABOR CHARGE, CORE Ordered / Back ordered : 0.5 / 0.0 EA			
0040	1 EA	MPINS INSURANCE AND HANDLING CHARGE Ordered / Back ordered : 1 / 0 EA			
0050	1 EA	FREIGHT FREIGHT TO DRAEGER Ordered / Back ordered : 1 / 0 EA			
0060	1 EA	FREIGHT FREIGHT TO AGENCY Ordered / Back ordered : 1 / 0 EA			