

# 1 4/4/12  
ARAJ 0075

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08754</b>		2. Contact: <b>Tpr. I Steve Dellanoce #6027</b>	4. Date: <b>03/23/2012</b>
		3. Phone Number: <b>609-947-9211</b>	5. County: <b>Ocean</b>
6. Alcotest Instrument Serial Number: <b>ARAJ-0075</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Complaint received of instrument experiencing control test failures.**

10. Comments:  
**Operational condition of the instrument was assessed by reviewing the performance history of the instrument. This revealed low Electrochemical Sensor (EC) results (Fuel Cell) on control tests. Instrument placed out of service and returned to Draeger Safety Diagnostics for evaluation.**  
**Please replace Fuel Cell.**

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Tpr. I Steve Dellanoce #6027**  
Name & Badge Number (Print)

**Tpr I Steve Dellanoce # 6027** **3/23/12**  
Signature Date

4/4/12  
ARAJ0075

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Twp Police Department 255 Oak Ave Toms River, NJ 08754</b>		2. Contact: <b>Tpr. I Steve Dellanoce #6027</b>	4. Date: <b>04/04/2012</b>
		3. Phone Number: <b>609-947-9211</b>	5. County: <b>Ocean</b>
6. Alcotest Instrument Serial Number: <b>ARAJ-0075</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**Alcotest instrument returned from outside maintenance and placed back in service.**

10. Comments:  
**See Draeger Return and Repair form.**

- 11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
- 12. The above Instrument/Component is placed out of service pending further evaluation.
- 13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

**Tpr. I Steve Dellanoce #6027**  
Name & Badge Number (Print)

*Tpr I Steve Dellanoce #6027*  
Signature

**4/4/12**  
Date

ARAJ 0075 4/4/12

### RETURN AND REPAIR FORM



#### Customer Information

B:

S:

Company Name: Toms River Twp, Toms River, NJ

Date Received: 03/27/12 Date given to service: 03/27/12

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR AJ-0075  
Printer Serial#: AR  
Sim Serial#: DD  
Probe Serial#: DD

Warranty Expires: \_\_\_\_\_

Description:  A  B  Plus  Demo  Screener  Trade In

#### Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

#### Repair Information:

Test#: \_\_\_\_\_

Part Number	Description	Qty	Total Cost
<u>mp Cal 71</u>	<u>Calibration</u>	<u>1</u>	<u>w/cw</u>
<u>ops</u>			
<u>6808455</u>	<u>Fuel cell</u>	<u>1</u>	<u>w/cw</u>
<u>mp labor</u>	<u>labor</u>	<u>.5</u>	<u>w/cw</u>

Repair Notes: Replaced Fuel cell # ARAE-0124 w/ Fuel Cell # ~~ops~~ ARCM-0307.

cal w/ops; ops check

Service Technician: [Signature] Date: 3-29-12

4/4/12  
ARAJ 0075

# Packing Slip

Customer no    Order No    Order date  
150059761    10539371    03/29/2012

Packing slip no.    Ship date  
80674980    03/29/2012

Please reference on inquiries

ship to  
150059763  
TOWNSHIP OF TOMS RIVER POLICE DEPAR  
ATTN: CHIEF OF POLICE  
OCEAN COUNTY  
255 OAK AVENUE  
TOMS RIVER, NJ 08754  
USA

bill to

Your Purchase Order Number

7110: ARAJ-0075  
Any questions? Please contact:  
MARA HERRERA

Customer  
TOWNSHIP OF TOMS RIVER  
OCEAN COUNTY  
255 OAK AVE  
TOMS RIVER, NJ 08753-0876  
USA

Your sales person

Don Pouliot  
Phone: 973-398-3228

Delivery terms

FA  
FREIGHT ALLOWED

Page 1 / 2

Item#	sh. Quant.	Part no. Description			
		Ship via: Fed Ex Saver			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARAJ-0075			
0010	1 EA	MPCAL71 CALIBRATION CHARGE - 7110/9510 Ordered / Back ordered : 0 / 0 EA			
0020	1 EA	6808455 SENSOR : FUEL CELL Ordered / Back ordered : 0 / 0 EA			

**Draeger Safety Diagnostics Inc.**  
Accounting Address:  
101 Technology Drive  
Pittsburgh, PA 15275  
Tel: (412) 788-5537  
Fax: (412) 788-5598

**Remit Wire Transfers To:**  
Citizens Bank  
Acct. # 6209426615  
Acct. Name: Draeger Safety Diagnostics Inc.  
Transit Number: 036076150  
Federal ID Number: 84-1600159

**REMIT TO:**  
Draeger Safety Diagnostics Inc.  
P.O. Box 200337  
Pittsburgh, PA 15251-0337

4/4/12  
AR17J0075

# Packing Slip

Customer no    Order No    Order date  
150059761    10539371    03/29/2012

Packing slip no.    Ship date  
80674980    03/29/2012

Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0030	0.5 H	MPLABOR LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 H			