



ARWF-0394  
11-4-16

STATE OF NEW JERSEY  
DEPARTMENT OF LAW AND PUBLIC SAFETY  
DIVISION OF STATE POLICE  
**BREATH TESTING INSTRUMENTATION SERVICE REPORT**

1. Department: <b>Toms River Police Department 255 Oak Ave Toms River, NJ 08753</b>		2. Contact: <b>Sgt. Robert Rankin</b>	4. Date: <b>11/4/2016</b>
		3. Phone Number: <b>732-349-0150</b>	5. County: <b>Ocean</b>
6. Alcotest Instrument Serial Number: <b>ARWF-0394</b>	7. Simulator Component Serial Number: <b>N/A</b>	8. Temperature Probe Component Serial Number: <b>N/A</b>	

9. Reason for Service:  
**The alcotest instrument was returned from outside evaluation and placed back in service.**

10. Comments:  
**See Draeger Return and Repair form.**

Last Known Sequential File #: N/A

11. The above Instrument/Component has been found to be in satisfactory working condition; no further action required.
12. The above Instrument/Component is placed out of service pending further evaluation.
13. The above Instrument/Component is placed back in service.

"I have been certified by the Attorney General as a Breath Test Coordinator/Instructor pursuant to N.J.A.C. 13:51, et. seq. In my official capacity as a Breath Test Coordinator/Instructor I perform inspections of evidential breath test devices and related components of evidential breath test systems. I have inspected the evidential breath test device listed on this report in my official capacity. The results of my inspection are recorded on this form. I certify that the forgoing statements made by me are true and accurate to the best of my knowledge. I am aware that if any of the forgoing statements made by me are willingly false, I am subject to punishment."

Tpr. I Kevin W. Alcott #6704  
Name & Badge Number (Print)

Tpr. I K W. A #6704 11/4/16  
Signature Date

# RETURN AND REPAIR FORM



## Customer Information

B:

S:

Company Name: Toms River P.D., Toms River, NJ

Date Received: 09/06/2016

Date given to service: 09/06/2016

Carrier:  FedEx  UPS  USPS Shipping Method:  GRD  3DAY  2DAY  
 NDA-PRI  NDA-STD

Product:  6510  6810  
 7110  8610  
 7510  DT5000  
 7410 Upper-half  9510  
 7410 Whole

Serial Number: AR WF-0394

Printer Serial#: AR

Sim Serial#: DD

Probe Serial#: DD

Warranty Expires: December 2016

Description:  A  B  Plus  Demo  Screener  Trade In

## Accessories

110V A/C Adapter  Regulator  Printer Ribbon  Printer Paper  
 Mouthpieces  9510 Stylus  9510 Top Cover  Carrying Case  
 Dry Gas  Other (please specify) \_\_\_\_\_

## Repair Information:

## Test#:

Part Number	Description	Qty	Total Cost
4414161	Calibration	1	NC/W
4414166	Labor	.5	NC/W

Repair Notes: Recalibrated unit, performed and passed all QC.

CAL W/QC AND OPS CHECK

Service Technician BC

Date: 10/11/2016

# Packing Slip

Customer no    Order No    Order date  
 150059761    10855215    10/11/2016

Packing slip no.    Ship date  
 81032668    10/11/2016

Please reference on inquiries

Ship to  
 150059763  
 TOWNSHIP OF TOMS RIVER POLICE DEPAR  
 ATTN: CHIEF OF POLICE  
 OCEAN COUNTY  
 255 OAK AVENUE  
 TOMS RIVER, NJ    08753-3383  
 USA

Payer

Your Purchase Order Number

SGT. ROBERT RANKIN  
 Any questions? Please contact:  
 JASMINE COLLINS

Customer  
 TOWNSHIP OF TOMS RIVER  
 OCEAN COUNTY  
 255 OAK AVE  
 TOMS RIVER, NJ    08753-0876  
 USA

Your sales person

CHRISTOPHER FAIRCHILD  
 Phone: 866-385-5900  
 Fax: 972-929-1260

Delivery terms  
 PPA  
 PREPAY AND ADD

Page    1 / 2

Number of packages: 1

Item#	sh. Quant.	Part no. Description			
		Waybill number: 9622001900001029848400677251655990 Ship via: Fed Ex Ground			
		NOTE: All claims for shortages/defects must be reported within 10 Business days after receipt of order. All returns must have our RMA# clearly marked on the box. Please call Customer Service at 866-385-5900.			
		7110: ARWF-0394 REC'D 9/6/2016			
0010	1 EA	4414161 CALIBRATION, 7110 Ordered / Back ordered : 0 / 0 EA			

**Draeger Safety Diagnostics Inc.**  
 Accounting Address:  
 101 Technology Drive  
 Pittsburgh, PA 15275  
 Tel: (412) 788-5537  
 Fax: (412) 788-5598

**Remit Wire Transfers To:**  
 Citizens Bank  
 Acct. # 6209426615  
 Acct. Name: Draeger Safety Diagnostics Inc.  
 Transit Number: 036076150  
 Federal ID Number: 84-1600159

**REMIT TO:**  
 Draeger Safety Diagnostics Inc.  
 P.O. Box 536410  
 Pittsburgh, PA 15253-5906

# Packing Slip

Customer no    Order No    Order date  
150059761    10855215    10/11/2016

Packing slip no.    Ship date  
81032668    10/11/2016  
Please reference on inquiries

Page    2 / 2

Item#	sh. Quant.	Part no. Description			
0020	0.5 EA	4414166 HOURLY LABOR CHARGE Ordered / Back ordered : 0.0 / 0.0 EA			