

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02255	Calib. Date: 01/30/2018
Certification File No.:	02256	Cert. No.: 00030
Linearity File No.:	02257	Cert. Date: 01/30/2018
Solution File No.:	02375	Lin. No.: 00021
Sequential File No.:	02375	Lin. Date: 01/30/2018
		Soln. No.: 00278
		Soln. Date: 06/14/2018
		File Date: 06/14/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDAN-0041
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 0827

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:51D		
Control 1 EC	0.100%	14:51D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:52D		
Control 2 EC	0.099%	14:53D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	14:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:53D		
Control 3 EC	0.099%	14:54D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:54D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:55D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Pascal Gambardella #365

Badge No.: 365

Date: 06/14/2018