

SOLN # 262

6/28/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01659
Certification File No.: 01660
Linearity File No.: 01661
Solution File No.: 01681
Sequential File No.: 01681

Calib. Date: 01/30/2018
Cert. Date: 01/30/2018
Lin. Date: 01/30/2018
Soln. Date: 06/28/2018
File Date: 06/28/2018

Serial No.: ARWF-0394
Calib. No.: 00039
Cert. No.: 00029
Lin. No.: 00029
Soln. No.: 00262

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17290

Model No.: CU-34

Serial No.: DDAN-0040
Expires: 09/14/2019
Bottle No.: 1358

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:26D		
Control 1 EC	0.100%	12:27D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:27D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27D		
Control 2 EC	0.100%	12:28D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:28D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:29D		
Control 3 EC	0.099%	12:29D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:30D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: KOEPPEN

First Name: ADAM

MI: M

Signature:  #350

Badge No.: 350

Date: 06/28/2018