

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02255	Calib. Date:	01/30/2018
Certification File No.:	02256	Cert. No.:	00030
Linearity File No.:	02257	Cert. Date:	01/30/2018
Solution File No.:	02394	Lin. No.:	00021
Sequential File No.:	02394	Lin. Date:	01/30/2018
		Soln. Date:	07/13/2018
		Soln. No.:	00280
		File Date:	07/13/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDAN-0041
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	1173

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:42D		
Control 1 EC	0.100%	14:43D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:43D		
Control 2 EC	0.098%	14:44D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:44D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:45D		
Control 3 EC	0.098%	14:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	14:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:46D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 07/13/2018