

SOLN 263

7-18-18

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.		
Calibration File No.:	01682	Calib. Date:	07/18/2018
Certification File No.:	01683	Cert. No.:	00040
Linearity File No.:	01684	Cert. Date:	07/18/2018
Solution File No.:	01685	Lin. No.:	00030
Sequential File No.:	01685	Soln. Date:	07/18/2018
		Soln. No.:	00263
		File Date:	07/18/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0387
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	1125

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:15D		
Control 1 EC	0.102%	12:16D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:16D		
Control 2 EC	0.101%	12:17D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	12:17D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:18D		
Control 3 EC	0.101%	12:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:19D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDWJ P2-026 (KA)

**Changed By:**

Last Name: ALCOTT First Name: KEVIN MI: W.

Signature: Sgt. K W. Q #6704

Badge No.: 6704  
Date: 07/18/2018