



ORIGINAL

Calibrating Unit

New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C

Serial No.: ARAJ-0075

Location:

TOMS RIVER P.D.

Calibration File No.: 02397

Calib. Date: 07/18/2018

Calib. No.: 00031

Certification File No.: 02398

Cert. Date: 07/18/2018

Cert. No.: 00022

Linearity File No.: 02399

Lin. Date: 07/18/2018

Lin. No.: 00022

Solution File No.: 02433

Soln. Date: 08/19/2018

Soln. No.: 00284

Sequential File No.: 02433

File Date: 08/19/2018

Calibrating Unit:

WET

Model No.: CU-34

Serial No.: DDWK S3-0386

Control Solution %: 0.100%

Expires: 12/11/2019

Solution Control Lot: 17420

Bottle No.: 1139

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	12:23D		
Control 1 EC	0.098%	12:24D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	12:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:25D		
Control 2 EC	0.097%	12:25D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	12:25D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:26D		
Control 3 EC	0.098%	12:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	12:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:27D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: WESTBALL

First Name: ROBERT

MI: S

Signature: _____

Badge No.: 306

Date: 08/19/2018