

SOLN 285
8/29/18

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02397	Calib. Date:	07/18/2018
Certification File No.:	02398	Cert. No.:	00031
Linearity File No.:	02399	Cert. Date:	07/18/2018
Solution File No.:	02448	Lin. No.:	00022
Sequential File No.:	02448	Lin. Date:	07/18/2018
		Soln. Date:	08/29/2018
		Soln. No.:	00285
		File Date:	08/29/2018
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	17420	Expires:	12/11/2019
		Bottle No.:	1142

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	18:03D		
Control 1 EC	0.101%	18:04D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	18:04D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:04D		
Control 2 EC	0.100%	18:05D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	18:05D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:06D		
Control 3 EC	0.099%	18:06D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	18:06D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	18:07D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA First Name: PASCAL MI: M

Signature:  #365

Badge No.: 365
Date: 08/29/2018