

SOLN # 266

8/30/18

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01682
Certification File No.: 01683
Linearity File No.: 01684
Solution File No.: 01698
Sequential File No.: 01698

Serial No.: ARWF-0394

Calib. Date: 07/18/2018
Cert. Date: 07/18/2018
Lin. Date: 07/18/2018
Soln. Date: 08/30/2018
File Date: 08/30/2018

Calib. No.: 00040
Cert. No.: 00030
Lin. No.: 00030
Soln. No.: 00266

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 17420

Model No.: CU-34

Serial No.: DDWK S3-0387
Expires: 12/11/2019
Bottle No.: 1167

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	19:27D		
Control 1 EC	0.101%	19:28D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	19:28D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:29D		
Control 2 EC	0.100%	19:29D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	19:29D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:30D		
Control 3 EC	0.101%	19:31D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	19:31D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:31D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 08/30/2018