

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02397	Calib. Date: 07/18/2018
Certification File No.:	02398	Calib. No.: 00031
Linearity File No.:	02399	Cert. Date: 07/18/2018
Solution File No.:	02483	Cert. No.: 00022
Sequential File No.:	02483	Lin. Date: 07/18/2018
		Lin. No.: 00022
		Soln. Date: 10/26/2018
		Soln. No.: 00290
		File Date: 10/26/2018
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	17420	Expires: 12/11/2019
		Bottle No.: 1154

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	12:07D		
Control 1 EC	0.102%	12:08D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	12:08D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:09D		
Control 2 EC	0.100%	12:09D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.102%	12:09D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:10D		
Control 3 EC	0.100%	12:10D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.102%	12:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	12:11D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature:  #365

Badge No.: 365

Date: 10/26/2018