SOLN 296
12/29/18

Calibrating Unit
New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C			Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.				
Calibration File No.:	02535	Calib. Date:	12/27/2018	Calib. No.:	00032
Certification File No.:	02536	Cert. Date:	12/27/2018	Cert. No.:	00023
Linearity File No.:	02537	Lin. Date:	12/27/2018	Lin. No.:	00023
Solution File No.:	02541	Soln. Date:	12/29/2018	Soln. No.:	00296
Sequential File No.:	02541	File Date:	12/29/2018		

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWK S3-0386
Control Solution %:	0.100%			Expires:	05/10/2020
Solution Control Lot:	18150			Bottle No.:	1367

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	15:08S		
Control 1 EC	0.095%	15:09S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.096%	15:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	15:10S		
Control 2 EC	0.094%	15:10S	33.9°C	CONTROL TEST FAILED
Control 2 IR	0.096%	15:10S	33.9°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:11S		
Control 3 EC	0.093%	15:12S	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.096%	15:12S	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	15:13S		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in acordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA First Name: PASCAL MI: M

Signature: #365 Badge No.: 365

Date: 12/29/2018