

# Calibrating Unit

## New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARWF-0394
Location:	TOMS-RIVER P.D.	
Calibration File No.:	01714	Calib. Date: 12/27/2018
Certification File No.:	01715	Calib. No.: 00041
Linearity File No.:	01716	Cert. Date: 12/27/2018
Solution File No.:	01731	Cert. No.: 00031
Sequential File No.:	01731	Lin. Date: 12/27/2018
		Lin. No.: 00031
		Soln. Date: 02/07/2019
		Soln. No.: 00275
		File Date: 02/07/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0387
Solution Control Lot:	18320	Expires: 09/26/2020
		Bottle No.: 1078

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	16:25S		
Control 1 EC	0.100%	16:26S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	16:26S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:27S		
Control 2 EC	0.098%	16:27S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	16:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:28S		
Control 3 EC	0.098%	16:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	16:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	16:30S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: \_\_\_\_\_

Badge No.: 365

Date: 02/07/2019