

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARWF-0394
Location:	TOMS RIVER P.D.	Calib. Date:	12/27/2018
Calibration File No.:	01714	Calib. No.:	00041
Certification File No.:	01715	Cert. Date:	12/27/2018
Linearity File No.:	01716	Lin. Date:	12/27/2018
Solution File No.:	01733	Soln. Date:	02/16/2019
Sequential File No.:	01733	File Date:	02/16/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDWK S3-0387
Control Solution %:	0.100%			Expires:	09/26/2020
Solution Control Lot:	18320			Bottle No.:	1088

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	10:13S		
Control 1 EC	0.100%	10:14S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	10:14S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:15S		
Control 2 EC	0.098%	10:16S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	10:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:16S		
Control 3 EC	0.098%	10:17S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	10:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	10:18S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: BAYER

First Name: RONALD

MI: C

Signature: _____



Badge No.: 1355

Date: 02/16/2019