

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARAJ-0075
Location:	TOMS RIVER P.D.	
Calibration File No.:	02535	Calib. Date: 12/27/2018
Certification File No.:	02536	Cert. No.: 00032
Linearity File No.:	02537	Cert. Date: 12/27/2018
Solution File No.:	02579	Lin. Date: 12/27/2018
Sequential File No.:	02579	Lin. No.: 00023
		Soln. Date: 02/22/2019
		Soln. No.: 00308
		File Date: 02/22/2019
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDWK S3-0386
Solution Control Lot:	18150	Expires: 05/10/2020
		Bottle No.: 1248

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	19:31S		
Control 1 EC	0.100%	19:32S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	19:32S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:33S		
Control 2 EC	0.098%	19:33S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	19:33S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:34S		
Control 3 EC	0.097%	19:35S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	19:35S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	19:35S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: GAMBARDELLA

First Name: PASCAL

MI: M

Signature: _____

Badge No.: 365

Date: 02/22/2019