

# Calibrating Unit

## New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01714 Calib. Date: 12/27/2018 Calib. No.: 00041  
Certification File No.: 01715 Cert. Date: 12/27/2018 Cert. No.: 00031  
Linearity File No.: 01716 Lin. Date: 12/27/2018 Lin. No.: 00031  
Solution File No.: 01739 Soln. Date: 03/14/2019 Soln. No.: 00278  
Sequential File No.: 01739 File Date: 03/14/2019

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDWK S3-0387  
Control Solution %: 0.100% Expires: 09/26/2020  
Solution Control Lot: 18320 Bottle No.: 1101

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:27S		
Control 1 EC	0.099%	22:27S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	22:27S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:28S		
Control 2 EC	0.098%	22:29S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	22:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:30S		
Control 3 EC	0.098%	22:30S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.099%	22:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:31S		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: DEROSA

First Name: MICHAEL

MI: A

Signature: 

Badge No.: 386

Date: 03/14/2019