

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARAJ-0075
Location:	TOMS RIVER P.D.		
Calibration File No.:	02535	Calib. Date:	12/27/2018
Certification File No.:	02536	Cert. No.:	00032
Linearity File No.:	02537	Cert. Date:	12/27/2018
Solution File No.:	02596	Lin. Date:	12/27/2018
Sequential File No.:	02596	Lin. No.:	00023
		Soln. Date:	03/14/2019
		Soln. No.:	00312
		File Date:	03/14/2019
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDWK S3-0386
Solution Control Lot:	18320	Expires:	09/26/2020
		Bottle No.:	0892

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	22:15S		
Control 1 EC	0.097%	22:16S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.097%	22:16S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:17S		
Control 2 EC	0.096%	22:17S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.096%	22:17S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:18S		
Control 3 EC	0.095%	22:19S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.098%	22:19S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:20S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


**Changed By:**

Last Name: DEROSA

First Name: MICHAEL

MI: A

Signature: \_\_\_\_\_



Badge No.: 386

Date: 03/14/2019