

Calibrating Unit New Standard Solution Report

ORIGINAL

Sol. #195
1/10/15

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01221
Certification File No.: 01222
Linearity File No.: 01223
Solution File No.: 01234
Sequential File No.: 01234

Calib. Date: 11/05/2014
Cert. Date: 11/05/2014
Lin. Date: 11/05/2014
Soln. Date: 01/10/2015
File Date: 01/10/2015

Serial No.: ARWF-0394
Calib. No.: 00028
Cert. No.: 00020
Lin. No.: 00020
Soln. No.: 00195

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 14B127

Model No.: CU-34

Serial No.: DDWK S3-0387
Expires: 02/14/2016
Bottle No.: 0818

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	14:04S		
Control 1 EC	0.101%	14:05S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	14:05S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:05S		
Control 2 EC	0.099%	14:06S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.099%	14:06S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:07S		
Control 3 EC	0.100%	14:08S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	14:08S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:08S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____



Badge No.: 1282

Date: 01/10/2015