

Calibrating Unit
New Standard Solution Report

ORIGINAL

SOL.#/68

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01521
Certification File No.: 01522
Linearity File No.: 01523
Solution File No.: 01589
Sequential File No.: 01589

Serial No.: ARAJ-0075

3-5-15

Calib. Date: 11/05/2014 Calib. No.: 00017
Cert. Date: 11/05/2014 Cert. No.: 00012
Lin. Date: 11/05/2014 Lin. No.: 00012
Soln. Date: 03/05/2015 Soln. No.: 00168
File Date: 03/05/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 14F130

Model No.: CU-34

Serial No.: DDWK S3-0386
Expires: 06/20/2016
Bottle No.: 0017

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	17:51S		
Control 1 EC	0.099%	17:52S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	17:52S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:52S		
Control 2 EC	0.097%	17:53S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.097%	17:53S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:54S		
Control 3 EC	0.097%	17:54S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	17:54S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	17:55S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____



Badge No.: 1282

Date: 03/05/2015