

# Calibrating Unit New Standard Solution Report

501# 198  
4/7/15

**Equipment** Alcotest 7110 MKIII-C Serial No.: ARWF-0394  
Location: TOMS RIVER P.D.  
Calibration File No.: 01254 Calib. Date: 04/07/2015 Calib. No.: 00029  
Certification File No.: 01255 Cert. Date: 04/07/2015 Cert. No.: 00021  
Linearity File No.: 01256 Lin. Date: 04/07/2015 Lin. No.: 00021  
Solution File No.: 01257 Soln. Date: 04/07/2015 Soln. No.: 00198  
Sequential File No.: 01257 File Date: 04/07/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040  
Control Solution %: 0.100% Expires: 02/14/2016  
Solution Control Lot: 14B127 Bottle No.: 0811

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:41D		
Control 1 EC	0.100%	11:42D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	11:42D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:43D		
Control 2 EC	0.099%	11:43D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	11:43D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:44D		
Control 3 EC	0.098%	11:45D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	11:45D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	11:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Temperature Probe Serial Number: DDXKP2-444

**Changed By:**

Last Name: DELLANOCE

First Name: JOSEPH

MI: S

Signature: SGT. DELLANOCE # 6027

Badge No.: 6027

Date: 04/07/2015