

ORIGINAL

Calibrating Unit New Standard Solution Report

501.299
4/22/15

Equipment Alcotest 7110 MKIII-C Serial No.: ARWF-0394
Location: TOMS RIVER P.D.
Calibration File No.: 01254 Calib. Date: 04/07/2015 Calib. No.: 00029
Certification File No.: 01255 Cert. Date: 04/07/2015 Cert. No.: 00021
Linearity File No.: 01256 Lin. Date: 04/07/2015 Lin. No.: 00021
Solution File No.: 01258 Soln. Date: 04/22/2015 Soln. No.: 00199
Sequential File No.: 01258 File Date: 04/22/2015

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDAN-0040
Control Solution %: 0.100% Expires: 06/20/2016
Solution Control Lot: 14F130 Bottle No.: 0025

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	14:50D		
Control 1 EC	0.099%	14:50D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.098%	14:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:51D		
Control 2 EC	0.097%	14:52D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	14:52D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:53D		
Control 3 EC	0.097%	14:53D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	14:53D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	14:54D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: _____



Badge No.: 1282

Date: 04/22/2015