

Calibrating Unit
New Standard Solution Report

ORIGINAL

Sol. # 201
5/13/15
Failed (W)

Equipment Alcotest 7110 MKIII-C
Location: TOMS RIVER P.D.
Calibration File No.: 01254
Certification File No.: 01255
Linearity File No.: 01256
Solution File No.: 01262
Sequential File No.: 01262

Serial No.: ARWF-0394

Calib. Date: 04/07/2015 Calib. No.: 00029
Cert. Date: 04/07/2015 Cert. No.: 00021
Lin. Date: 04/07/2015 Lin. No.: 00021
Soln. Date: 05/13/2015 Soln. No.: 00201
File Date: 05/13/2015

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 14B127

Model No.: CU-34 Serial No.: DDAN-0040
Expires: 02/14/2016
Bottle No.: 0817

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	11:53D		
Control 1 EC	0.093%	11:53D	34.0°C	CONTROL TEST FAILED
Control 1 IR	0.098%	11:53D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	11:54D		
Control 2 EC	0.092%	11:55D	34.0°C	CONTROL TEST FAILED
Control 2 IR	0.097%	11:55D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	11:56D		
Control 3 EC	0.092%	11:57D	34.0°C	CONTROL TEST FAILED
Control 3 IR	0.097%	11:57D	34.0°C	CONTROL TEST FAILED
Ambient Air Blank	0.000%	11:57D		

Test results are not within acceptable range.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: MAIRE

First Name: GUY

MI: E

Signature: *JEM #282*

Badge No.: 1282

Date: 05/13/2015